

Median Nerve Stimulation to Detect Awake and Anesthetized States with Riemannian Classifiers

Valérie Marissens Cueva

valerie.marissens@loria.fr

Supervisors:

Fabien Lotte, Laurent Bougrain, Sébastien Rimbart

Accidental Awareness during General Anesthesia (AAGA)

- **Unexpected awakening** during surgery under general anesthesia [1], which can cause pain and potential psychological sequelae [2].
- During AAGA, patients **try to move** to alert the anesthetist [3], yet it isn't always possible due to the **neuromuscular-blocking agents** (NMBAs) which paralyze them.



Heleen Blussé van Oud-Alblas

[1] Pandit et al. British Journal of Anaesthesia, 2014.

[2] Osterman et al. Gen Hosp Psychiatry. 2001.

[3] Sandin et al. The Lancet, 2000.

[4] Bruhn et al. Br. J. Anaesth, 2006.

[5] Myles et al. Lancet, 2004.

[6] Sebel et al. Anesth. Analg. 2004.


Accidental Awareness during General Anesthesia (AAGA)

- **Unexpected awakening** during surgery under general anesthesia [1], which can cause pain and potential psychological sequelae [2].
- During AAGA, patients **try to move** to alert the anesthetist [3], yet it isn't always possible due to the **neuromuscular-blocking agents** (NMBAs) which paralyze them.
- Traditional clinical signs are unreliable indicators of anesthesia depth [4], which leads to the use of **brain activity monitors**.
- Awareness **still occurs** with current monitors [5, 6].

[1] Pandit et al. British Journal of Anaesthesia, 2014.

[2] Osterman et al. Gen Hosp Psychiatry. 2001.

[3] Sandin et al. The Lancet, 2000.



Bispectral Index monitor, Sigl and Chamoun, J Clin Monit, 1994

[4] Bruhn et al. Br. J. Anaesth, 2006.

[5] Myles et al. Lancet, 2004.

[6] Sebel et al. Anesth. Analg. 2004.

Long-term objective

Propose a new Brain-Computer Interface (BCI) based on motor imagery to:

- monitor the motor cortex with EEG
- detect movement intentions
- and alert the medical staff [7]

feasible with current BCIs, although the MI vs rest classification remains low [8] for a reliable device to be used in a hospital

Median Nerve Stimulation (MNS)-based BCI

- **Painless stimulation** of the median nerve, with an intensity between 8 and 15 mA.




Illustration of the regular branching of the median nerve [9]




Illustration of ERD/ERS induced by MI, MNS and MNS during MI, in the alpha and beta frequency bands [7].

[7] Rimbert et al. Front. Neurosci, 2019.

[9] Lanz, J. of Hand Surgery, 1977.

Median Nerve Stimulation (MNS)-based BCI

- **Painless stimulation** of the median nerve, with an intensity between 8 and 15 mA.
- Combining motor intention (MI) and MNS significantly impacts MNS-generated patterns, leading to **better MI detection accuracy** (+12%) in patients under light sedation [7].





Illustration of the regular branching of the median nerve [9]



Distribution of average classification accuracies (n = 16), obtained by a TS+LR classifier, ***p-value < 0.001 [7].

Median Nerve Stimulation (MNS)-based BCI

- **Painless stimulation** of the median nerve, with an intensity between 8 and 15 mA.
- Combining motor intention (MI) and MNS significantly impacts MNS-generated patterns, leading to **better MI detection accuracy** (+12%) in patients under light sedation [7].
- Provides a **trigger**.





Illustration of the regular branching of the median nerve [9]



Distribution of average classification accuracies (n = 16), obtained by a TS+LR classifier, ***p-value < 0.001 [7].

[7] Rimbert et al. Front. Neurosci, 2019.

[9] Lanz, J. of Hand Surgery, 1977.

Problematic

- The MNS pattern is still visible under light sedation [10, 11].
- But are we still able to detect it at deeper concentrations of anesthesia?


Is it possible to use median nerve stimulation to differentiate states of consciousness under general anesthesia?

Clinical protocol at CHU Brugmann

10 volunteers: 6 females; 50 ± 8.62 years old




[12] Rimbart et al. JMIR Res. Protoc, 2023.




Data collection: protocol

MNS: Median Nerve Stimulation



- 1 run = 150 stimulations spaced by 3 to 4 seconds
- the duration and number of runs depended on the length of the surgery

Impact of GA on the patterns induced by MNS



Grand average time frequency analysis across 8 subjects for electrode C4


MNS sensorimotor modulations [13] are strongly **influenced by general anesthesia**: PSR, ERD and PMBR disappear

[13] Rimbart et al. 2th IEEE Intern. Winter Conf. on BCI, 2024.


Riemannian Minimum Distance to the Mean (MDM)




MNS-awake




MNS-anesthesia




Riemannian Minimum Distance to the Mean (MDM)




Riemannian Minimum Distance to the Mean (MDM)




Offline classification: MNS-awake vs MNS-anesthesia




Offline classification: ~~MNS~~-awake vs ~~MNS~~-anesthesia




Test set balanced accuracy of the 10 subjects.
The yellow line indicates the average test set balanced accuracy (97%).





Distance between test covariance matrices and the MNS-awake class centroid



Ideal distances of each test trial to the MNS-awake centroid




Distance between test covariance matrices and the ~~MNS~~ awake class centroid



Distances of each test trial to the MNS-awake centroid, for subjects 6 and 8.

Distance between test covariance matrices and the ~~MNS~~ awake class centroid



Distances of each test trial to the MNS-awake centroid, for subject 9

- **External elements** polluted the signal (repositioning electrodes, the operating table, intubation...)
- **Age of 75** vs average of 50 years old [14]
- **Fewer training trials** (although no correlation between the number of trials and accuracies)

[16] Purdon et al. Br. J. Anaesth, 2015



Detect Awake and Anesthetized states following the Median Nerve Stimulation pattern with a Riemannian MDM classifier?



- Correctly identifies the periods when the patient is awake
- **Covariance matrices associated to the MNS pattern seem to evolve consistently with the patient's level of awareness.**



- In a real-world application, **deep anesthesia trials will not be available** to calibrate the BCI. Thus, a **one class approach** will be required.



One-Class Riemannian EEG Classifier to Detect Anesthesia

Valérie Marissens Cueva^{1,2}, Sébastien Rimbert², Ana Maria Cebolla Alvarez³, Mathieu Petieau³, Iraj Hashemi³, Viktoriya Vitkova³, Guy Cheron³, Claude Meistelman⁴, Philippe Guerci⁴, Denis Schartz⁵, Seyed Javad Bidgoli⁶, Laurent Bougrain², Fabien Lotte¹

Acknowledgments



Fabien Lotte
Research Director
Inria Center at the University of Bordeaux



Laurent Bougrain
Associate Professor
Université de Lorraine



Sébastien Rimbart
Inria Starting Faculty Position
Inria Center at the University of Bordeaux

Collaborators:

- Ana Maria Cebolla Alvarez
- Mathieu Petieau
- Viktoriya Vitkova
- Iraj Hashemi
- Guy Cheron
- Claude Meistelman
- Denis Schmartz
- Philippe Guerçi
- Seyed Javad Bidgoli






Thank you for your attention!

Contact: valerie.marissens@loria.fr

Informations des patients



Interventions

Fistule anale	40 min
Cholecystectomie	40 min
Annexectomie coelio Gauche	1h15
Sphinteroplastie anale	1h20
Coelioscopie pour endométriose	2h10
TURP	1h40
Hemmoroides	40 min
Fermeture d'Ileostomie	1h30
Changement de plaque orthopédique	2h20

Monitors to prevent AAGA




Narcotrend monitor

www.narcotrend.de



SedLine monitor

www.maasimo.fr




Bispectral Index monitor, Sigl and Chamoun, J Clin Monit, 1994

www.nursekey.com

Future analysis

- **Other EEG features** (entropy, functional connectivity, somatosensory evoked potentials) will be investigated
- Identification of the **specific frequency band and most relevant electrodes** to detect changes in the MNS pattern when a patient is anesthetized
- Validation: a protocol will be carefully conceived to try to **simulate an AAGA** and observe if the BCI will indeed be able to detect it.

Impact of GA on the patterns induced by MNS




Grand average time frequency analysis across 8 subjects for electrode C4

- Electrical artifact is also modified: other possible factors are changes in skin conductance, environment, patient's arm position between the two sessions

Offline classification: ~~MNS~~-awake vs ~~MNS~~-anesthesia

- Classifier: Riemannian Minimum Distance to the Mean
- 64 electrodes
- Preprocessing :
 - downsampling 128 Hz
 - 8-30 Hz band pass filter
 - epochs : 250 to 1000 ms

Electrocautery-related EEG artifacts



ANR BCI4IA protocols

MOTANA:

- CHRU de Nancy
- 0, 0.5 and 1 $\mu\text{g/ml}$ propofol
- healthy volunteers

STIM-MOTANA:

- CHU de Brugmann
- before and during surgery under GA





Diagram of the experiment (adapted from S. Rimbert et al., Trials, 2019)

Illustration of ERD/ERS induced by MI, MNS and MNS during MI, in the alpha and beta frequency bands [4]



[4] S. Rimbert et al. Front. Neurosci, 2019.

[5] Schnitzler et al. Neuroimage, 1997.

Riemannian Minimum Distance to the Mean (MDM)

- Each EEG trial $\mathbf{X}$ is represented by a **covariance matrix** $\mathbf{P} \in \mathbb{R}^{n \times n}$, with n the number of electrodes
- Estimation of the **centroids** $\mathbf{P}_{\text{MNS-awake}}$ and $\mathbf{P}_{\text{MNS-anesthesia}}$, the mean covariance matrices of each class
- The class $\hat{k}$ of an unseen EEG covariance matrix $\mathbf{P}_x$ is determined based on the **nearest centroid's class**:

$$\hat{k} = \operatorname{argmin}_i (\delta_R(\mathbf{P}_x, \mathbf{P}_i))$$

with δ_R the distance between two matrices

[9] M. Congedo et al. 8th Inter. BCI Conf. 2019.

[10] A. Barachant et al. IEEE Trans. Biomed. Eng. 2012.

States of consciousness under general anesthesia

awake	anesthetized	
conscious	conscious	unconscious